



ESTANDAR CORROSION TEST FOR LPG BY COOPER STRIP TEST ASTM D 1838/07

- Built from 304/316L Stainless Steel
- External Surface with industrial polish
- Hydraulic pressure test at 1000 PSI
- Leakage test at 500 PSI
- Two SS-316 needle valves
- Two cooper strips
- Optional: D-130 Adjunct Comparative table for cooper strip
- On demand certificates:
 - Quality of the Raw Material of the Cylinder and valves
 - Hydraulic test of the cylinder and valves
 - Leak test of the unit
 - Quality of Raw Material of the cooper strip

