



Pozos 1335 – Lomas del Mirador - CP B1752EOA / [www.gascontrol.com.ar](http://www.gascontrol.com.ar)  
Teléfono: 54 4652-1128 – 2091-3732 / [info@gascontrol.com.ar](mailto:info@gascontrol.com.ar)

## STANDARD TEST METHOD FOR CORROSIVNESS TO COOPER FROM PETROLUM PRODUCTS BY COOPER STRIP TEST

- Built from 304 / 316L Stainless steel.
- Threaded system with O ring seal. Hydrocarbon suitable
- External Surface with industrial polish
- Hydraulic pressure test at 100 PSI
- Internal holder for cooper strip
- External holder for thermostatic Bath
- Two cooper Strip
- Does not includes glass tube.
- On demand certificates:
  - Quality of the Raw Material of the Cylinder
  - Hydraulic test
  - Leak test
  - Quality of Raw Material of the cooper strip

